

MEASUREMENT EXPLORES~ INSIGHT & EFFICIENCY



GDM-906X Series

6 1/2 Digit Dual Measurement Multimeter

- 6 1/2 Digit Display: 1,200,000 Counts
- 4.3" TFT Graphic LCD
- DCV Basic Accuracy: 0.0035%(GDM-9061)/0.0075%(GDM-9060)
- 12 Measurement Functions: DCV, ACV, DCI, ACI, 2-wire and 4-wire Resistance, Frequency, Period, Diode, Continuity, Temperature and Capacitance
- Sampling Rate up to 10k SPS (GDM-9061)
- Dual Measurements to Perform Two Selected Measurement Simultaneously
- Offer Graphical Capabilities Including Histogram, Bar Meter and Trend
- Temperature Measurement Support RTD, Thermistor as Well as Thermocouple
- Standard Interface: USB Host/Device, RS-232C, LAN, Digital I/O
- Optional Interface: GPIB



SPECIFICATIONS

DC CHARACTERISTICS Accuracy : ± (% of reading + % of range)

DC Voltage				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Input Resistance			
100.0000 mV	0.1μV	10MΩ or >10GΩ	GDM-9061	GDM-9060	
1.000000 V	1μV	10MΩ or >10GΩ	0.0050 + 0.0035	0.0090 + 0.0065	
10.00000 V	10μV	10MΩ or >10GΩ	0.0048 + 0.0007	0.0080 + 0.0010	
100.0000 V	0.1mV	10MΩ ±1%	0.0035 + 0.0005	0.0075 + 0.0005	
1000.000 V	1mV	10MΩ ±1%	0.0050 + 0.0006	0.0085 + 0.0006	
			0.0050 + 0.0010	0.0085 + 0.0010	

Resistance				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Test Current			
100.0000 Ω	100μΩ	1mA	GDM-9061	GDM-9060	
1.000000 kΩ	1mΩ	1mA	0.010 + 0.004	0.014 + 0.007	
10.00000 kΩ	10mΩ	100μA	0.010 + 0.001	0.014 + 0.001	
100.0000 kΩ	100mΩ	10μA	0.010 + 0.001	0.014 + 0.001	
1.000000 MΩ	1Ω	5μA	0.010 + 0.001	0.014 + 0.001	
10.00000 MΩ	10Ω	500nA	0.040 + 0.001	0.040 + 0.001	
100.0000 MΩ	100Ω	500nA//10 MΩ	0.800 + 0.010	0.800 + 0.010	

DC Current				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Burden Volt.			
100.0000 μA	100pA	< 0.011 V	GDM-9061	GDM-9060	
1.000000 mA	1nA	< 0.11 V	0.05 + 0.025	0.05 + 0.025	
10.00000 mA	10nA	< 0.04 V	0.05 + 0.006	0.05 + 0.006	
100.0000 mA	100nA	< 0.4 V	0.05 + 0.020	0.05 + 0.020	
1.000000 A	1μA	< 0.7 V	0.05 + 0.005	0.05 + 0.005	
3.000000 A	1μA	< 2.0 V	0.10 + 0.010	0.10 + 0.010	
10.00000 A	10μA	< 0.5 V	0.20 + 0.020	0.20 + 0.020	
			0.15 + 0.010	-----	

Continuity				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Test Current			
1000.000 Ω	0.001 Ω	1 mA	GDM-9061	GDM-9060	
			0.010 + 0.030	0.014 + 0.030	

Diode Test				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Test Current			
5.000000 V	1μV	1 mA	GDM-9061	GDM-9060	
			0.010 + 0.030	0.014 + 0.030	

DC Ratio
Accuracy Specification: ± (DC Input accuracy + DC Reference accuracy)

TEMPERATURE CHARACTERISTICS

RTD (Accuracy based on PT100)

Range	Resolution	Accuracy(1Year)(TCAL±5°C)
-200 °C ~ -100 °C	0.001 °C	0.09 °C
-100 °C ~ -20 °C	0.001 °C	0.08 °C
-20 °C ~ 20 °C	0.001 °C	0.06 °C
20 °C ~ 100 °C	0.001 °C	0.08 °C
100 °C ~ 300 °C	0.001 °C	0.12 °C
300 °C ~ 600 °C	0.001 °C	0.22 °C

Thermocouples (Accuracy based on ITS-90)

Type	Range	Resolution	Accuracy(1Year)(TCAL±5°C)
E	-200 °C ~ +1000 °C	0.002 °C	0.2 °C
J	-210 °C ~ +1200 °C	0.002 °C	0.2 °C
T	-200 °C ~ +400 °C	0.002 °C	0.3 °C
K	-200 °C ~ +1372 °C	0.002 °C	0.3 °C
N	-200 °C ~ +1300 °C	0.003 °C	0.4 °C
R	-50 °C ~ +1768 °C	0.01 °C	1 °C
S	-50 °C ~ +1768 °C	0.01 °C	1 °C
B	+350 °C ~ +1820 °C	0.01 °C	1 °C

Thermistor (2.2kΩ, 5kΩ, 10kΩ or User Type)

Range	Resolution	Accuracy(1Year)(TCAL±5°C)
-80 °C ~ 150 °C	0.01 °C	0.01 °C

ORDERING INFORMATION

GDM-9061	6 1/2 (1200000 counts) Digit Dual Measurement Multimeter
GDM-9060	6 1/2 (1200000 counts) Digit Dual Measurement Multimeter
ACCESSORIES	
Safety Instructions x 1, Power cord x 1, USB cable GTL-246 x 1, Test lead GTL-207A x 1, CD x1 (including the complete user manual, upgrade program and PC software, DMM-Viewer2)	

Global Headquarters
GOOD WILL INSTRUMENT CO., LTD.
 No.7-1, Jhongsing Road, Tucheng Dist., New Taipei City 236, Taiwan
 T +886-2-2268-0389 F +886-2-2268-0639
 E-mail: marketing@goodwill.com.tw

China Subsidiary
GOOD WILL INSTRUMENT (SUZHOU) CO., LTD.
 No. 521, Zhuijiang Road, Snd, Suzhou Jiangsu 215011 China
 T +86-512-6661-7177 F +86-512-6661-7277
 E-mail: marketing@instek.com.cn

Malaysia Subsidiary
GOOD WILL INSTRUMENT (SEA) SDN. BHD.
 No. 1-3-18, Elit Avenue, Jalan Mayang Pasir 3, 11950 Bayan Baru, Penang, Malaysia
 T +604-6111122 F +604-6115225
 E-mail: sales@goodwill.com.my

Europe Subsidiary
GOOD WILL INSTRUMENT EURO B.V.
 De Run 5427A, 5504DG Veldhoven, THE NETHERLANDS
 T +31 (0)40-2557790 F +31 (0)40-2541194

U.S.A. Subsidiary
INSTEK AMERICA CORP.
 5198 Brooks Street Montclair, CA 91763, U.S.A.
 T +1-909-399-3535 F +1-909-399-0819
 E-mail: sales@instekamerica.com

Japan Subsidiary
TEXIO TECHNOLOGY CORPORATION.
 7F Towa Fudosan Shin Yokohama Bldg., 2-18-13 Shin Yokohama, Kohoku-ku, Yokohama, Kanagawa, 222-0033 Japan
 T +81-45-620-2305 F +81-45-534-7181
 E-mail: info@texio.co.jp

Korea Subsidiary
GOOD WILL INSTRUMENT KOREA CO., LTD.
 Room No.503, Gyeonginro 775 (Mullae-Dong 3Ga, Ace Hightech-City B/D 1Dong), Yeongduengpo-Gu, Seoul 150093, Korea.
 T +82-2-3439-2205 F +82-2-3439-2207
 E-mail: gwinstek@gwinstek.co.kr

AC CHARACTERISTICS Accuracy : ± (% of reading + % of range)

AC Voltage (True RMS)				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Frequency			
100.0000 mV	0.1μV	3Hz ~ 5Hz	GDM-9061	GDM-9060	
			1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
			0.06 + 0.04	0.09 + 0.04	
			0.12 + 0.05	0.15 + 0.05	
			0.60 + 0.08	0.63 + 0.08	
1.000000 V to 750.000 V	1μV ~ 1mV	3Hz ~ 5Hz	1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
			0.06 + 0.04	0.09 + 0.04	
			0.12 + 0.05	0.15 + 0.05	
			0.60 + 0.08	0.63 + 0.08	
			4.00 + 0.50	4.00 + 0.50	

AC Current (True RMS)				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution	Frequency			
100.0000 μA to 10.00000 mA	100pA to 10nA	3Hz ~ 5Hz	GDM-9061	GDM-9060	
			1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
			0.10 + 0.04	0.13 + 0.04	
			0.18 + 0.04	0.20 + 0.04	
			1.00 + 0.04	1.00 + 0.04	
1.000000 mA to 100.0000 mA	1nA to 100nA	3Hz ~ 5Hz	1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
			0.10 + 0.04	0.13 + 0.04	
			0.15 + 0.04	0.18 + 0.04	
			1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
1.000000 A	1μA	3Hz ~ 5Hz	1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
			0.10 + 0.04	0.13 + 0.04	
			0.15 + 0.04	0.18 + 0.04	
			1.00 + 0.04	1.00 + 0.04	
			0.35 + 0.04	0.38 + 0.04	
3.000000 A	1μA	3Hz ~ 5Hz	1.10 + 0.04	-----	
			0.35 + 0.04	-----	
			0.23 + 0.04	0.23 + 0.04	
			0.23 + 0.04	0.23 + 0.04	
			1.10 + 0.04	-----	
			0.35 + 0.04	-----	
10.00000 A	10μA	3Hz ~ 5Hz	1.10 + 0.04	-----	
			0.35 + 0.04	-----	
			0.15 + 0.04	-----	
			0.35 + 0.04	-----	
			0.15 + 0.04	-----	
			0.35 + 0.04	-----	

CAPACITANCE CHARACTERISTICS Accuracy : ± (% of reading + % of range)

Capacitance				Accuracy(1Year)(TCAL±5°C)	
Range	Resolution				
1.000 nF	0.001nF	2.00 + 2.00			
10.00 nF	0.01nF	2.00 + 1.00			
100.0 nF	0.1nF	2.00 + 0.40			
1.000 μF	0.001μF	2.00 + 0.40			
10.00 μF	0.01μF	2.00 + 0.40			
100.0 μF	0.1μF	2.00 + 0.40			

FREQUENCY AND PERIOD CHARACTERISTICS Accuracy : ± (% of reading)

Frequency/Period				Accuracy(1Year)(TCAL±5°C)	
Range	Frequency				
100.0000mV to 750.000V	3Hz ~ 5Hz	0.1			
		0.05			
		0.03			
		0.006			

GENERAL INFORMATION

Display	4.3" Color TFT WQVGA (480 x 272)
Standard Interface	RS-232C, USB Host/Device, LAN, Digital I/O
Power Source	AC 100 V/120 V/220 V/240 V±10%
Power Line Frequency	50 Hz/60 Hz/400 Hz±10%
Power Consumption	Max. 25VA
Dimension & Weight	267(W) x 107(H) x 302(D) mm, Approx. 3.5kg

Specifications subject to change without notice. GDM-906XCD1DH

OPTION

Opt.1 GPIB card	(*) GPIB can be installed at customer site
-----------------	--

OPTIONAL ACCESSORIES

GTL-205	Temperature Probe Adapter with Thermal Coupling (K-type), approx. 1000mm	GRA-422	Rack Mount Kit(19",2U)
GTL-234	RS-232C Cable, 9-pin female-female cable, approx. 2000mm		
GTL-248	GPIB Cable, approx. 2000mm		
GTL-308	4Wire Type (+shield) Test lead, approx. 1500mm		



www.gwinstek.com

www.facebook.com/GWInstek